

CERTIFICATE

OF BIOCOMPATIBILITY

Material tested: remanium® GM 900
Manufacturer: DENTAURUM J.P. Winkelstroeter KG
Turnstraße 31 D-75228 Ispringen

Test:
The following test was carried out in accordance with ISO 10993-5, 1992: „Test on cytotoxicity; *in vitro* methods“. The test samples were manufactured by DENTAURUM.

CYTOTOXICITY

The cytotoxic effect of the test material was examined *in vitro* by the cell culture test using L 929-cells (fibroblast cell line); Examination of materials for medical devices with the extraction method and protein determination.

Result: remanium® GM 900 showed no cytotoxic effects.

BIOSERVICE Scientific Laboratories GmbH Munich

Behringstraße 6
D-82152 Planegg



Dr. Albrecht Poth
Date: February 03, 2000

